

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100690

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:06/03/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.48: 0.079
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 1059 psi

BLANK TEST	0.000	06:01
INTERNAL STANDARD	VERIFIED	06:01
EXTERNAL STANDARD	0.076	06:01
BLANK TEST	0.000	06:02
EXTERNAL STANDARD	0.076	06:02
BLANK TEST	0.000	06:03
EXTERNAL STANDARD	INTERFERENCE	06:04

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.0°C	PASSED

PUMP INFO

Flow Rate = 4.340 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only JMW 06/14/21