

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 4/7/16 Instrument Serial Number: 100414
Instrument Location: EMK AK 10140899
Supervisor Name and Agency: NEMUS AST

Describe the Status Change or Issue with the Instrument:

OUT OF SERVICE

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 2/5/2017 1300
4/7/16
CWF 4/8/16

For Use by SCDL Additional Notes

Last uploaded subject test - 12/13/15
Tank pressure too low for VOC - 1/3/16
Tank pressure too low for VOC - 2/2/16
Tank changed / VOC Passed - 3/29/16
Called Nick Massey on 4/5/16 to confirm correct out of service date - CWF 4/5/16

Removed from service for preventative maintenance. CWS 6/14/16
Email completed form to ALASKA.SCDL@AKSTATELABORATORY.COM

For questions contact the Breath Alcohol Program at 907-269-5740