

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100384

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:09/08/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 687 psi

BLANK TEST	0.000	23:45
INTERNAL STANDARD	VERIFIED	23:45
EXTERNAL STANDARD	0.075	23:45
BLANK TEST	0.000	23:46
EXTERNAL STANDARD	0.075	23:46
BLANK TEST	0.000	23:47
EXTERNAL STANDARD	0.075	23:47
BLANK TEST	0.000	23:48
EXTERNAL STANDARD	0.074	23:48

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.3°C	PASSED

PUMP INFO

Flow Rate = 4.738 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------