

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100689

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/30/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.90: 0.077
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 871 psi

BLANK TEST	0.000	14:32
INTERNAL STANDARD	VERIFIED	14:33
EXTERNAL STANDARD	0.076	14:33
BLANK TEST	0.000	14:34
EXTERNAL STANDARD	0.075	14:34
BLANK TEST	0.000	14:35
EXTERNAL STANDARD	0.075	14:35
BLANK TEST	0.000	14:36
EXTERNAL STANDARD	0.075	14:36
BLANK TEST	0.000	14:37
EXTERNAL STANDARD	0.075	14:37
BLANK TEST	0.000	14:38

Average = 0.0752
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.07
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 45.2°C PASSED

PUMP INFO

Flow Rate = 3.943 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 11/10/22

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10 day of 01, 2022



Notary Public
My Commission Expires With Office



00312/7/21