

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/21/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.12: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 871 psi

BLANK TEST	0.000	12:50
INTERNAL STANDARD	VERIFIED	12:50
EXTERNAL STANDARD	0.079	12:50
BLANK TEST	0.000	12:51
EXTERNAL STANDARD	0.076	12:51
BLANK TEST	0.000	12:52
EXTERNAL STANDARD	INTERFERENCE	12:52

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 45.5°C	PASSED

PUMP INFO

Flow Rate = 4.410 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only
02-24-25