

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100398

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/18/2013

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.18: 0.080  
LOT #: 02012080A1  
EXPIRATION: 03/01/2014

BLANK TEST	0.000	13:19
INTERNAL STANDARD	VERIFIED	13:19
EXTERNAL STANDARD	0.079	13:19
BLANK TEST	0.000	13:20
EXTERNAL STANDARD	0.079	13:20
BLANK TEST	0.000	13:21
EXTERNAL STANDARD	0.079	13:21
BLANK TEST	0.000	13:22
EXTERNAL STANDARD	0.079	13:22
BLANK TEST	0.000	13:23
EXTERNAL STANDARD	0.079	13:23
BLANK TEST	0.000	13:24

Average = 0.0790  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 2.00  
PIC: 3.02  
Modem: 2.3  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 46.6°C PASSED

### PUMP INFO

Flow Rate = 5.332 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD


PASSED



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 8/2/2013  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Aug. 2013

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CRF  
7/24/13