

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100361

Received From: APD

Date: 7/26/24

Analyst: C. O. Bryant

Reason for instrument return:

Plastic end of breath hose where the mouth piece is attached was removed + thrown away. New breath hose needed.

Change instrument and ^{005 7/26/24} ~~case~~ location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)