

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100675

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 07/18/2025

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C      PASSED  
Breath Tube      = 44.6°C      PASSED

PUMP INFO  
Flow Rate = 4.561 L/M      PASSED

DETECTOR INFO  
PUMP ON      PASSED  
PUMP OFF      PASSED

FILTER INFO  
Filter 1      PASSED  
Filter 2      PASSED  
Filter 3      PASSED

INTERNAL STANDARD      FAILED

Diagnostic Check Failed

File Only  
CW 8/15/25