

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100686

Received From: Whittier PD

Date: 4/27/22

Analyst: Colleen O'Bryant

Reason for instrument return:

DMT returned to compare DMHost data + records to DMT long term storage data + records due to the inability to upload via modem while in the field.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

* Keyboard was not included and the external standard was still in the instrument. COB 5/6/22