

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 07/27/2018 Instrument Serial Number: 100350

Instrument Location: Glennallen AST Post

Supervisor Name and Agency: Trooper A. Garcia AST

Describe the Status Change or Issue with the Instrument:

Datamaster is out of service and being sent to the crime lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 07/27/2018 / 1830

-----  
**For Use by SCDL**

**Additional Notes**

Detector overflow on VIC - 7/15/2018 JMY 07/30/2018

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740