

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100360

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/26/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.65: 0.079
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 1199 psi

BLANK TEST	0.000	12:23
INTERNAL STANDARD	VERIFIED	12:23
EXTERNAL STANDARD	0.077	12:24
BLANK TEST	0.000	12:25
EXTERNAL STANDARD	0.077	12:25
BLANK TEST	0.000	12:26
EXTERNAL STANDARD	0.077	12:26
BLANK TEST	0.000	12:27
EXTERNAL STANDARD	0.076	12:27
BLANK TEST	0.000	12:28
EXTERNAL STANDARD	0.077	12:28
BLANK TEST	0.000	12:29

Average = 0.0768
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 43.3°C PASSED

PUMP INFO

Flow Rate = 4.754 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD


PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 03 day of 12, 20 21




Notary Public
My Commission Expires With Office

066-11/8/21