

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 10-25-2016 Instrument Serial Number: 100347

Instrument Location: Cordova PD

Supervisor Name and Agency: Nate Taylor - Cordova PD

Describe the Status Change or Issue with the Instrument:

~~Several~~ Several "Gas Flow Errors" in the past month.
Usually only occurs on first test. Subsequent tests run fine.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10-25-2016 18:22 hours

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740