

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 10/13/15 Instrument Serial Number: 100417

Instrument Location: Wildwood Pretrial Facility - Kenai

Supervisor Name and Agency: Ryan Grimm - Kenai P.D.

Describe the Status Change or Issue with the Instrument:

Touchscreen does not work, was unable to perform a V.O.C.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 10/13/15 0340

For Use by SCDL

Additional Notes

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740