

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100414

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/30/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.67: 0.077  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 977 psi

BLANK TEST                    0.000 13:00  
INTERNAL STANDARD        VERIFIED 13:00

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C        PASSED  
Breath Tube        = 48.1°C        PASSED

### PUMP INFO

Flow Rate = 4.692 L/M        PASSED

### DETECTOR INFO

PUMP ON                        PASSED  
PUMP OFF                       PASSED

### FILTER INFO

Filter 1                        PASSED  
Filter 2                        PASSED  
Filter 3                        PASSED

### INTERNAL STANDARD

PASSED