

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 06/06/18 Instrument Serial Number: 100368

Instrument Location: Nancy Lake Ranger Station, Willow Alaska

Supervisor Name and Agency: W. Stuart Kirksey / Alaska State Troopers

Describe the Status Change or Issue with the Instrument:
Return to Service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 06/06/2018 / 0430 hours

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

*Previous tank leaked down to empty. New o-rings provided
to agency. JMV 06/06/18*

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740