

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100381

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 05/15/2025

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.50: 0.079  
LOT #: AG436502  
EXPIRATION: 12/30/2026  
TANK PRESSURE: 1074 psi

|                   |              |       |
|-------------------|--------------|-------|
| BLANK TEST        | 0.000        | 22:45 |
| INTERNAL STANDARD | VERIFIED     | 22:45 |
| EXTERNAL STANDARD | INTERFERENCE | 22:45 |

Interference Detected

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

|                         |        |
|-------------------------|--------|
| Sample Chamber = 48.9°C | PASSED |
| Breath Tube = 44.4°C    | PASSED |

|                       |        |
|-----------------------|--------|
| PUMP INFO             |        |
| Flow Rate = 4.514 L/M | PASSED |

|               |        |
|---------------|--------|
| DETECTOR INFO |        |
| PUMP ON       | PASSED |
| PUMP OFF      | PASSED |

|             |        |
|-------------|--------|
| FILTER INFO |        |
| Filter 1    | PASSED |
| Filter 2    | PASSED |
| Filter 3    | PASSED |

|                   |        |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|

File Only JMW 06/18/25