

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/12/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.02: 0.078

LOT #: AG436502

EXPIRATION: 12/30/2026

TANK PRESSURE: 1208 psi

BLANK TEST 0.000 15:10

INTERNAL STANDARD VERIFIED 15:11

EXTERNAL STANDARD INTERFERENCE 15:11

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED

Breath Tube = 44.5°C PASSED

PUMP INFO

Flow Rate = 4.549 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only DMT 06/26/25