

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 08/05/20 Instrument Serial Number: 100359

Instrument Location: DETACH AST POST

Supervisor Name and Agency: E. JOHNSON AST

Describe the Status Change or Issue with the Instrument:

COULD PSR LEAK VOC

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 08/05/20 1608 Hours

For Use by SCDL

Additional Notes

Failed VOC due to pump issue JMW 08/10/20

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100359

Date: 08/03/2020
Time: 14:33:04

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 46.5°C PASSED

PUMP INFO

Flow Rate = 4.491 L/M PASSED

DETECTOR INFO

PUMP ON FAILED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100359

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:08/03/2020

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 47.4°C PASSED

PUMP INFO

Flow Rate = 4.426 L/M PASSED

DETECTOR INFO

PUMP ON FAILED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

Diagnostic Check Failed