

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/31/2023

Instrument #: 100368

Instrument Location: Seldovia Police Department

Supervisor Name and Agency: Tatiana Malnoske, ASCDL

Describe the Status Change or Issue with the Instrument:

Instrument taken out of service as of 05-15-2023. On 07-24-2023, I spoke to VPSO Jared Patterson of Seldovia PD who advised me that there are currently no breath test operators and supervisors at their agency, and that the instrument still has tank AG113402 installed which expired on 05-14-2023. The DMT has been turned off since no one is able to switch the tanks out.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 05-15-2023

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor