

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/20/19 Instrument Serial Number: 100374

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack Anchorage Police

Describe the Status Change or Issue with the Instrument:

Per crime lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10/10/19 2052 hours

For Use by SCDL

Additional Notes

Intermittent elevated filter readings.
See In field Instrument Review Jm 10/21/19

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740