

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100399

Received From Intoximeters

Date 02/03/21

Analyst Derek walton

1. Reason for instrument return:

Refurbished by manufacturer, repaired for failing to
upload via modem line

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

Instrument was not in
the field
DMV 02/03/21

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

SRO Number: SRO-043153

Received Date: 11/25/2020

Completion Date: 01/08/2021

SRO Type: REPAIR

SRO Description: DMT Preventative Maintenance

Customer ID: C000003651

Customer PO: 112520-Barnett

Contact: Derek Walton

Email: derek.walton@alaska.gov

Ship To:

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1	**All Instruments are Calibrated to Factory Specifications**
Unit Type:	(LEGDMT) LEGACY DMT
Serial Number:	50100399
Reason:	Preventive Maintenance
	planned
Reason Notes:	For Preventive Maintenance.
Resolution:	Update
	DMTVR3
Resolution Notes:	Installed TB-30 DMT Digital Pots Mod on Control Board. Cleaned & adjusted filter wheel solenoid. Updated filter wheel solenoid mounting screws. Updated stepper motor adjustment lock down screw. Replaced filter wheel motor pinion. Cleaned, lubricated, & adjusted filter wheel moving parts. Removed, inspected, and applied metallic seal tape to mirror blocks of Sample Chamber. Inspected 5-Way Valve. Optimized detector position. Adjusted Tech Screen settings. Performed Calibration/Adjustment, Wet and Dry Gas Accuracy Checks, Breath Tests, and outgoing Diagnostics Test.

Parts:	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	90-8000-00	CONTRACT PREV MAINT DMT
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN

Order Packing Slip Report

Intoximeters
Experience • Service • Integrity

1/14/2021 2:10:45 PM

Packing Slip: 219058

Pack Date: 1/14/2021

SRO: SRO-043153



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 2

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507
USA

Customer PO: 112520-Barnett

Ship Via: FDX

FED EX

Weight: 37.00

Delivery Terms : 2D

FOB : ORIGIN

Packages: 1

Customer Ship Account :

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100399
1	90-8000-00	EA	1.00	1.00	0.00
CONTRACT PREV MAINT DMT					
2	001199-00	EA	1.00	1.00	0.00
DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN					

Rec'd 02/03/21
JM