

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100418

Received From Kenai PD

Date 03/03/20

Analyst Derck Watten

1. Reason for instrument return:

Internal standard errors

2. Change instrument ~~and case~~ location in JusticeTrax Breath Alcohol Database

Not returned with case DMV 03/03/20

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)