In Field Instrument Review

Date Range: 2024.07.19 - 2024.09.05

Review Completed on 2024.10.07 by Kaija Swisher

Status Messages and Action Taken

100344 (Juneau PD): 1/1 Diagnostic Failed on VOCs due to pump failure. Passing VOC occurred on 9/18/24 – No action taken.

100351 (AST/Soldotna): 7/29 RFI on Subject Tests with same operator, 1/29 Pump Error on Subject Tests followed by passing subject test, operator was contacted regarding RFI messages – <u>100351</u> 2024.10.04 Change in Status.

100354 (Anchorage PD): 1/127 RFI on Subject Tests, 1/127 Pump Error on Subject Tests, 2/127 Interference Detected on Subject Tests, 1/127 Ambient Fail on Subject Tests. RFI, Pump Error, and Ambient Fail were all followed by passing subject tests – No action taken.

100358 (AST/Mat-Su West): 3/12 RFI on Subject Tests followed by passing subject tests – No action taken.

100365 (Skagway PD): 2/9 Internal Standard Errors on Subject Tests followed by passing subject tests. Due to upcoming preventative maintenance, instrument was removed from service – <u>100365</u> 2024.09.27 Change In Status.

100369 (AST/Bethel): 1/3 Tank Interference on VOCs, 1/3 Diagnostic Failed due to internal standard on VOCs. Taken out of service for preventative maintenance see <u>100369 2024.07.31 Change In</u> <u>Status.</u>

100371 (Whittier PD): 1/3 RFI on Subject Tests. Monitor for RFI Status Messages. *Recommend continued monitoring*.

100389 (AST/Fairbanks): 1/44 Pump Error on Subject Tests followed by passing subject tests – No action taken.

100392 (Homer PD): 2/16 RFI on Subject Tests, 1/16 Blank Error on Subject Tests, RFIs and Blank Error were both followed by completed subject tests – No action taken.

100400 (Bethel PD): 1/14 Interference on Subject Tests, 2/14 Tank Interference on Subject Tests, 1/14 Standard Out of Range on Subject Tests, 1/14 Ambient Fail on Subject Tests, all followed by completed subject tests – Recommend monitoring for Standard Out of Range and Tank Interference.

100402 (AST/Tok): 1/3 RFI on Subject Tests followed by passing subject test - No action taken.

100403 (Wasilla PD): 2/47 Interference Detected on Subject Tests, 1/47 Internal Standard Error on Subject Tests, all followed by passing subject tests – No action taken.

100413 (Fairbanks PD): 1/24 Pump Error on Subject Tests followed by passing subject tests – No action taken.

100414 (Glacier Bay National Park): 1/1 Diagnostic Failed on VOC due to internal standard error on 9/5, passing VOC for this period occurred on 9/17/24 – No action taken.

100415 (Anchorage PD): 1/11 Interference Detected on Subject Tests followed by passing subject tests – No action taken.

100420 (Ft. Greely): 2/5 RFI on VOCs followed by passing VOC, 1/5 Filter Wheel Error on VOCs followed by passing VOC, 1/4 Pump Error on Subject Tests "Test/Test" followed by passing test – Recommend monitoring for filter wheel errors.

100424 (Eielson AFB): 1/1 Diagnostic Failed on VOCs due to Pump Error, passing VOCs occurred on 9/18 – No action taken.

100668 (North Pole PD): 1/15 RFI on Subject Tests "Test/Test" - No action taken.

100669 (Palmer PD): 5/83 RFI on Subject Tests followed by passing subject tests, 4/83 Pump Errors on Subject Tests, 2/83 Interference Detected on Subject Tests – No action taken.

100672 (Wrangell PD): 1/6 Tank Interference on Subject Tests followed by passing subject test – No action taken.

100682 (Bristol Bay Borough PD): 1/2 Blank Error on Subject Tests followed by passing subject test – No action taken.

100689 (Dillingham PD): 1/7 RFI on Subject Tests followed by passing subject test – No action taken.

100691 (Cordova PD): 1/6 RFI on Subject Tests, 1/6 Interference Detected on Subject Tests followed by passing subject tests – No action taken.

100697 (Anchorage PD): 1/89 RFI on Subject Tests, 2/89 Blank Error on Subject Tests, 1/89 Standard Out of Range on Subject Tests, all tests followed by passing subject tests – No action taken.

100699 (Joint Base Elmendorf-Richardson): 3/25 RFI on Subject Tests, 1/25 Pump Error on Subject Tests followed by passing subject tests, 2/25 Internal Standard Error on Subject Tests followed by passing subject test, instrument removed from service – <u>100699 2024.10.03 Change in Status</u>.

Follow Up from Previous In Field Instrument Review

100365 (Skagway PD): 2/9 Internal Standard Errors on Subject Test followed by passing subject tests. Due to upcoming preventative maintenance, instrument was removed from service – <u>100365</u> 2024.09.27 Change In Status.

100371 (Whittier PD): 1/3 RFI on Subject tests. Monitor for RFI Status Messages. *Recommend continued monitoring*.

100382 (Petersburg PD): No status messages for this cycle – Discontinue monitoring for Static Analysis Errors.

100421 (Joint Base Elmendorf-Richardson): No status messages for this cycle – Discontinue monitoring for Blank Errors.

Agencies not included in this summary

100356 (Whittier PD - Girdwood): 1/1 Diagnostic Check failed due to pump failure on VOCs, Data retrieved from SD card at lab on 10/2/24, instrument returned to lab by agency on 9/27/24 – <u>100356</u> 2024.09.05 Change in Status.

100376 (AST/St. Mary's): Does not upload and no response on request for files. VOC from 07/18/2024 and 09/05/2024 are needed, VOC from 05/30/2024 is on file.

100381 (AST/Talkeetna): Instrument cannot currently upload, datafiles not sent at this time. Last VOC on file is from 06/04/2024. No breath test records since agency received instrument. Instrument was not able to upload as of 02/22/2024. - Recommend removing from service due to lack of use and connectivity.

100675 (AST/Angoon): Instrument cannot currently run VOCs due to Internal Standard error on VOC from 07/18/2024. Last VOC upload 05/30/2024. No breath test records since agency received instrument. There is currently no agency breath supervisor or operator at this location.

In Field Instrument Review Worksheet

Subject Test Data											Range		Total Tests	726								
		SL	lbjeci	t i es	st Dat	a				7/19/2024	9/5/2024	Tot	al Messages	62	8.5%							
Serial Number	Total Subject Tests		RFI	Pum	p Error Interference			Tank Interference		Blank Error		Standard Out of Range		Gas Flow Error		Ambient Fail		Internal Standard Error		Diagnostic Failed		Total Codes
100351	29	7	24%	1	3%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	8
100354	127	1	1%	1	1%	2	2%	0	0%	0	0%	0	0%	0	0%	1	1%	0	0%	0	0%	5
100358	12	3	25%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	3
100365	9	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2	22%	0	0%	2
100371	3	1	33%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100389	44	0	0%	1	2%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100392	16	2	13%	0	0%	0	0%	0	0%	1	6%	0	0%	0	0%	0	0%	0	0%	0	0%	3
100400	14	0	0%	0	0%	1	7%	2	14%	0	0%	1	7%	0	0%	1	7%	0	0%	0	0%	5
100402	3	1	33%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100403	47	0	0%	0	0%	2	4%	0	0%	0	0%	0	0%	0	0%	0	0%	1	2%	0	0%	3
100413	24	0	0%	1	4%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100415	11	0	0%	0	0%	1	9%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100420	4	0	0%	1	25%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100668	15	1	7%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100669	83	5	6%	4	5%	2	2%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	11
100672	6	0	0%	0	0%	0	0%	1	17%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100682	2	0	0%	0	0%	0	0%	0	0%	1	50%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100689	7	1	14%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100691	6	1	17%	0	0%	1	17%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2
100697	89	1	1%	0	0%	0	0%	0	0%	2	2%	1	1%	0	0%	0	0%	0	0%	0	0%	4
100699	25	3	12%	1	4%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2	8%	0	0%	6
Total	726	27	4%	10	1%	9	1%	3	0%	4	1%	2	0%	0	0%	2	0%	5	1%	0	0%	62

 $\frac{21}{21}$ + $\frac{35}{21}$ + \frac

In Field Instrument Review Worksheet

Effective:	1/4/2022
------------	----------

			t Test		D		5 /4	heel Error	7/19/2024	Range 9/5/2024		Total Tests al Messages	0	0.0%	0		Decently T				Table
Serial Number	Total Subject Tests	User Abort							Suck Back Error		Undefined Error		Filter 1 Won't Zero			namb Low		ube High		nalysis Error	
100342	3	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100344	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100346	9	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100347	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100351	29	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100354	127	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100358	12	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100361	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100362	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100364	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100365	9	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100366	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100367	7	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100370	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100371	3	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100374	5	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100378	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100379	4	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100380	11	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100382	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100388	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100389	44	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100390	23	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100392	16	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100398	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100400	14	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100402	3	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100403	47	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100404	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100405	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100412	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100413	24	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100415	11	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100417	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100418	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100420	4	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100665	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100667	7	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100668 100669	15 83	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100669	83 6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100672	15	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100673	15	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
	2	0 0%	0	0%	0		0	0%	0		0		0	0%	0	0%	0		0	0%	0
100680 100681	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100681	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100682	2	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100684	7	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100689	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100691	6	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100692	11	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100694	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100697	89	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100697	1	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100698	25	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100699	25	0 0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
Total	726	0 0%	Ő	0%	Ő	0%	Ő	0%	Ő	0%	Ö	0%	Ő	0%	0	0%	Ő	0%	Ő	0%	0
10101		0/0			, v						, v	0.0	, v	0.0		1 0/0	Ň	0.0	, v		, v

The date range listed above represents the first and last dates of data generated from the DMHost report. The search range could be larger than the date range listed.

In Field Instrument Review Worksheet

			VO	C Da	ata					Date 7/19/2024	Range 9/5/2024	Tot	Total Tests al Messages	95 7	7.4%						
Serial Number	Total VOC Tests		RFI		Pump Error		Interference		Tank Interference		Blank Error		Standard Out of Range		Gas Flow Error		nt Fail	Internal S	tandard Error	Diagnostic Faile	d Total Codes
100344	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1 100%	1
100369	3	0	0%	0	0%	0	0%	1	33%	0	0%	0	0%	0	0%	0	0%	0	0%	1 33%	2
100414	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1 100%	1
100420	5	2	40%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0 0%	2
100424	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1 100%	1
Total	95	2	2%	0	0%	0	0%	1	1%	0	0%	0	0%	0	0%	0	0%	0	0%	4 4%	7
		Th	e date ra	inge list	ted abov	e represe	ents the fi	rst and last	dates of	data genera	ted from the	DMHost re	port. The se	arch rang	e could be	e larger th	an the d	late range	listed.		

2024.07.19 - 2024.09.05 Page 6 of 7

	7/19/2024	9/5/2024		I Messages		1.1%																
Serial Number	Total VOC Tests	Detector	Overflow	Barome	eter Error	Filter W	Filter Wheel Error		Suck Back Error		Undefined Error		Filter 1 Won't Zero		Samp Chamb Low Breath Tube High			Static An	Total Codes			
100420	5	0	0%	0	0%	0	0%	1	20%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
Total	95	0	0%	0	0%	0	0%	1	1%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
		**	The date r	ange listed	above rep	resents the	e first and I	ast dates o	of data gen	erated from	the DMHost	report. Th	ne search r	ange could	d be larger	than the d	ate range li	sted.**				