

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100676

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/03/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.33: 0.073
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 1229 psi

BLANK TEST	0.000	15:13
INTERNAL STANDARD	VERIFIED	15:13
EXTERNAL STANDARD	0.073	15:13
BLANK TEST	0.000	15:14
EXTERNAL STANDARD	0.073	15:14
BLANK TEST	0.000	15:15
EXTERNAL STANDARD	0.073	15:15
BLANK TEST	0.000	15:16
EXTERNAL STANDARD	0.073	15:16
BLANK TEST	0.000	15:17
EXTERNAL STANDARD	0.073	15:17
BLANK TEST	0.000	15:18

Average = 0.0730
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 3.828 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

7/19/19

Subscribed and sworn before me this 19 day of 07, 20 19


Notary Public
My Commission Expires With Office



Jm 06/25/19