

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/19/2014

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.06: 0.080  
LOT #: 11313080A1  
EXPIRATION: 06/01/2015

BLANK TEST	0.000	19:41
INTERNAL STANDARD	VERIFIED	19:41
EXTERNAL STANDARD	0.080	19:41
BLANK TEST	0.000	19:42
EXTERNAL STANDARD	0.080	19:43
BLANK TEST	0.000	19:43
EXTERNAL STANDARD	0.080	19:44
BLANK TEST	0.000	19:45
EXTERNAL STANDARD	0.081	19:45
BLANK TEST	0.000	19:46
EXTERNAL STANDARD	0.081	19:46
BLANK TEST	0.000	19:47

Average = 0.0804  
Std Dev = 0.0005

## Diagnostic Check

VERSIONS  
DMT: 2.00  
PIC: 3.02  
Modem: 2.3  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C PASSED  
Breath Tube = 44.6°C PASSED

### PUMP INFO

Flow Rate = 5.371 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 4/8/2014  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 8<sup>th</sup> day of April, 20 14

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CRF  
4/1/14