

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 08/02/2024

Instrument #: B-100 G40

Instrument Location: UNALASKA, ALASKA ; JAIL BOOKING ROOM

Supervisor Name and Agency: GRANT DOWTY, UNALASKA DPS

Describe the Status Change or Issue with the Instrument:

· SCREEN ON DATAMASTER KEPT FREEZING, WOULD HAVE TO RESTART DATAMASTER.
· DATAMASTER WAS NOT RECOGNIZING THE PHONE LINE,
· PREVENTATIVE MAINTENANCE NEXT MONTH

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 08/02/2024 , 1032 HOURS

For Use by SCDL

Additional Notes

Empty box for additional notes.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor