

In Field Instrument Review

Dates: 09/05/2017 – 10/24/2017, Prepared by Colleen O’Bryant

100344 (King Cove PD) – 1/2 Ambient Fail – No action taken
100345 (AST/Soldotna) – 17/47 RFI, Interference, Tank Interference, Blank Error and Internal Standard Error – Instrument Removed from Service
100346 (Fairbanks Airport) – 3/8 – Pump Error, Interference, Blank Error – No action taken
100347 (Craig PD) – 1/5 – Gas Flow Error – No action taken
100350 (AST/Soldotna) – 8/11 – RFI, Tank Interference, Ambient Fail, Internal Standard Error – Instrument Removed from Service
100352 (Anchorage Airport PD) – 3/9 – Interference, all from the same subject – no action taken
100355 (Seward PD) – 1/7 – RFI – no action taken
100364 (APD) – 3/112 – RFI, Pump Error, Blank Error – no action taken
100367 (Palmer PD) – 2/2 – RFI, Filter Wheel Error – Instrument Removed from Service
100376 (Metlakatla PD) – 2/3 – Tank Interference - Instrument Removed from Service
100378 (AST/Fairbanks) – 6/52 – RFI, Gas Flow Error – tank changed, no further issues
100380 (Kenai PD) – 1/13 – Interference - no action taken
100384 (Juneau PD) – 3/24- RFI, Blank Error, Standard Out of Range – possibly environmental, same subject - no action taken
100385 (AST/Ketchikan) 1/5 – RFI - no action taken
100389 (Cordova PD) - 1/10 – Standard out of range – followed by a completed test - no action taken
100391 (Wrangell PD) - 11/21 – Pump Error, Tank Interference - Instrument Removed from Service
100396 (APD) – 12/127 – RFI, Pump Error, Interference, Blank Error, Ambient Fail, Internal Standard Error – will monitor instrument
100406 (Palmer PD) – 4/40 – RFI – no action taken
100414 (Eielson AFB) – 1/8 – Gas Flow Error – Done during “Test,Test”, no issues following - no action taken
100416 (Fairbanks PD) - 3/42 – RFI - no action taken
100421 (JBER) – 20/51 – RFI, Interference, Blank Error, Diagnostic Failed, Static Analysis Error - Instrument Removed from Service
100422 (JBER) - 3/12 – RFI – All done during “Test, Test” – no action taken
100669 (APD) – 2/29 – Interference – no action taken
100679 (Haines PD) - 1/2 - RFI – No action taken
100685 (AST/Tok) - 1/4 - RFI – No action taken
100692 (AST/Aniak) – 6/9 – Tank Interference, Standard out of Range, Ambient Fail – all conducted on the same subject – will monitor instrument
100695 (Wasilla PD) – 3/30 – RFI – no action taken
100696 (Soldotna PD) – 1/17 – Gas Flow Error – no action taken
100698 (Kenai PD) – 5/8 – Tank Interference, Standard out of Range – Instrument Removed from Service

