

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/05/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.96: 0.080
LOT #: 01315080A3
EXPIRATION: 02/05/2017
TANK PRESSURE: 729 psi

BLANK TEST	0.000	12:02
INTERNAL STANDARD	VERIFIED	12:02
EXTERNAL STANDARD	0.080	12:02
BLANK TEST	0.000	12:03
EXTERNAL STANDARD	0.080	12:03
BLANK TEST	0.000	12:04
EXTERNAL STANDARD	0.080	12:04
BLANK TEST	0.000	12:05
EXTERNAL STANDARD	0.079	12:05
BLANK TEST	0.000	12:06
EXTERNAL STANDARD	0.080	12:06
BLANK TEST	0.000	12:07

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.01
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 45.2°C PASSED

PUMP INFO

Flow Rate = 5.295 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

9/11/2015

Subscribed and sworn before me this 11th day of Sept., 2015


Nikki Roth, Notary Public
My Commission Expires With Office



BMB
9/10/15