

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: JUNE 27TH 2017 Instrument Serial Number: 100364
Instrument Location: ANCHORAGE JAIL BOOKING
Supervisor Name and Agency: SGT. RYAN ROCKOM APD

Describe the Status Change or Issue with the Instrument:

~ OUT OF SERVICE - BREATH TUBE WILL NOT HEAT UP AND DISPLAY
SCREEN WORKS ONLY INTERMITTENTLY

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____
Out of Service Date/Time: JUNE 25TH 2017 @ 0400 / OFFICER KINSLAND APD

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740