

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100364

Received From: Anchorage Airport

Date: 11/14/2023 2022  
DJW 11/15/2022

Analyst: Derek Walton

Reason for instrument return:

Breath tube temperature too low

Change instrument <sup>Not returned with case DJW 11/14/22</sup> and ~~case location~~ in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

### From Intoximeters

Intoximeters Repair Paperwork (attached)