

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100402

Received From APD

Date 10/10/18

Analyst Colleen O'Bryant

1. Reason for instrument return:

Possible 5 way valve issue.
Instrument not registering flow during
subject sample

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100402

Date: 09/28/2018
Time: 05:30:26

SUBJECT NAME:
TEST TEST
LICENSE #: 1234456

OPERATOR NAME:
DEREK W SITZ
OPERATOR'S #: 4096

DEPT/AGENCY:
ANCI
CASE/REPORT: 00-00000
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.15: 0.081
LOT #: AG735001
EXPIRATION: 12/16/2019

DIAGNOSTIC CHECK	PASSED	05:31
BLANK TEST	0.000	05:31
INTERNAL STANDARD	VERIFIED	05:31
EXTERNAL STANDARD	0.077	05:31
BLANK TEST	0.000	05:32
SUBJECT SAMPLE	INCOMPLETE	05:35

