

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100675

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 02/28/2025

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C

PASSED

Breath Tube = 48.1°C

PASSED

### PUMP INFO

Flow Rate = 4.850 L/M

PASSED

### DETECTOR INFO

PUMP ON

FAILED

PUMP OFF

FAILED

User Abort

File only  
0006-  
3/6/25