

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/14/21 Instrument Serial Number: 100699

Instrument Location: Wildwood Pretrial

Supervisor Name and Agency: M. Swengel - Kenai P.D.

Describe the Status Change or Issue with the Instrument:

Out of Service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10/14/21 @ 0130hrs

For Use by SCDL

Additional Notes

Removed from service due to internal standard errors on subject tests. Reference IFIR 2021.07.19 - 2021.09.05

COB 2021.10.19

Email completed form to dus.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740